



Australian Register of Therapeutic Goods Certificate

Issued to

Aero Healthcare

for approval to supply

Aero Healthcare - Wound hydrogel dressing, sterile

ARTG Identifier	221543
ARTG Start Date	21/03/2014
Product Category	Medical Device Included Class IIa
GMDN	47764
GMDN Term	Wound hydrogel dressing, sterile
Intended Purpose	A sterile wound covering impregnated with gel for first aid treatment of minor burns. The device is intended to cool the burn and soothe pain. It also acts a mechanical barrier to help protect the wound from infection. It may also absorb exudates in wounds that are minimally to moderately exudative. This is a single-use device.

Manufacturer Details	Address	Certificate number(s)
Aero Healthcare Aust Pty Ltd T/A Aero Healthcare	63 Seaton Street ARMIDALE , NSW , 2 350 Australia	DV-20081028-MC-062240-11

ARTG Standard Conditions

The above Medical Device Included Class IIa has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Wound hydrogel dressing, sterile

Product Specific Conditions

- No specific conditions have been recorded against this entry.