

two year limited warranty

Your Drive brand product is warranted to be free of defects in materials and workmanship for two years from the date of the original consumer purchase.

This device was built to exacting standards and carefully inspected prior to shipment. This Two Year Limited Warranty is an expression of our confidence in the materials and workmanship of our products and our assurance to the consumer of dependable service.

In the event of a defect covered by this warranty, we will, at our option, repair or replace the device.

This warranty does not cover device failure due to owner misuse or negligence, or normal wear and tear. The warranty does not extend to non-durable components such as rubber accessories, casters, and grips, which are subject to normal wear and need periodic replacement.

If you have a question about your Drive device or this warranty, please contact an authorized Drive dealer.



*Wenzel*iteRe/hab

© 2006 Medical Depot, Inc. All rights reserved.
Drive is a trademark of Medical Depot, Inc.
Port Washington N.Y. 11050 USA Made in China

otter tub stand

item # OT 8010

drive
MEDICAL DESIGN & MANUFACTURING

*Wenzel*iteRe/hab

www.drivemedical.com

otter tub stand
item # OT 8010

recommended use

The Otter Tub Stand elevates the Otter Bath Chair in the bathtub. The tub stand legs are adjustable and bath chair mounts easily on stand. All adjustments are tool free.

assembly

Your Otter Tub Stand comes in one carton completely assembled.
If your shipment is incomplete, please call a Drive authorized dealer.

maintenance and care

1. Inspect the tub stand periodically for cracks, breaks, loose or missing parts or malfunctions. If a product needs service or replacement parts, contact an authorized Drive dealer.
2. Wash tub stand with soap and water or 10% bleach.

adjustments

See Figure 2

1. Mount chair onto base by pushing down firmly on chair until legs snap into sockets (A) of tub stand. To remove chair from stand, depress red levers (B) one at a time and pull chair upwards.
2. To adjust legs (C), pull crossbar (D) outward, rotate legs and push in to engage.




figure 2