



Australian Register of Therapeutic Goods Certificate

Issued to

Aero Healthcare

for approval to supply

Aero Healthcare - Exudate-absorbent dressing, non-gel

ARTG Identifier	365072
ARTG Start Date	10/05/2021
Product Category	Medical Device Included Class IIa
GMDN	44970
GMDN Term	Exudate-absorbent dressing, non-gel
Intended Purpose	A sterile dressing intended to assist in wound healing by absorbing exudates in wounds and create a moist wound healing environment. The dressing is suitable to be applied to superficial wounds with moderate to high exudates including bedsores, trauma wounds, superficial burns, and other similar wounds. It does not produce a water-absorbent gel upon contact with wound exudate and does not contain an anti-microbial agent.

Manufacturer Details	Address	Certificate number(s)
Aero Healthcare Aust Pty Ltd T/A Aero Healthcare	63 Seaton Street ARMIDALE , NSW , 2 350 Australia	DV-2019-MC-00080-1

ARTG Standard Conditions

The above Medical Device Included Class IIa has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Exudate-absorbent dressing, non-gel

Product Specific Conditions

No specific conditions have been recorded against this entry.