

Isocitrate Assay Kit (Colorimetric)

LS-K218-100 (100 Tests) • Store at -20°C



Introduction

ISOCITRATE (ISOCITRIC ACID) is a substrate in the citric acid (TCA) cycle. Isocitrate is formed by the isomerization of citrate catalyzed by the enzyme aconitase. Isocitrate is oxidized by isocitrate dehydrogenase producing α -ketoglutarate and generating NADPH. Isocitrate is commonly found in many fruits and vegetables and their processed products. Industrially, isocitrate is used as a marker to identify the quality and purity of fruit juices. LSBio's isocitrate assay measures the NADPH generated from the oxidation of isocitrate. The NADPH converts the dye to an intense violet color with an absorption maximum at 565 nm. The increase in absorbance at 565 nm is directly proportional to the isocitrate concentration.

Key Features

- Fast and sensitive. Linear detection range (20 μ L sample): 20 to 5000 μ M for 10 min reaction.
- Convenient and high-throughput. Homogeneous "mix-incubate-measure" type assay. Can be readily automated on HTS liquid handling systems for processing thousands of samples per day.

Applications

- Isocitrate determination in food, beverage, biological samples (e.g. cell lysate, tissue homogenate, serum, etc.)

Components

Component	K218-100
	100 Tests
Assay Buffer	10 mL
Enzyme A	120 μ L
Enzyme B	120 μ L
NADP/MTT	1 mL
Standard	1 mL

Materials Not Supplied

Pipetting devices and accessories (e.g. multi-channel pipette), clear flat bottom 96-well plates, centrifuge tubes and plate reader.

Storage

The kit is shipped on ice. Store all kit components at -20 °C.

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121
www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com

Isocitrate Assay Kit (Colorimetric)

LS-K218-100 (100 Tests) • Store at -20°C



Assay Procedure

Sample Preparation

Tissue: Prior to dissection, rinse tissue in phosphate buffered saline (pH 7.4) to remove blood. Homogenize tissue (50 mg) in ~200 µL buffer containing 50 mM potassium phosphate (pH 7.5). Centrifuge at 10,000 x g for 15 min at 4°C. Remove supernatant for assay.

Cell Lysate: Collect cells by centrifugation at 2,000 x g for 5 min at 4°C. For adherent cells, do not harvest cells using proteolytic enzymes; rather use a rubber policeman. Homogenize or sonicate cells in an appropriate volume of cold buffer containing 50 mM potassium phosphate (pH 7.5). Centrifuge at 14,000 x g for 10 min at 4°C. Remove supernatant for assay.

All samples can be stored at -20 to -80°C for at least one month.

Reagent Preparation

Keep thawed Enzyme A and B on ice and equilibrate all other reagents to 25°C. Briefly centrifuge tubes before use.

Procedure using 96-well plate

- Standards. Prepare 200 µL 5000 µM Premix by mixing 10 µL of the Standard (100 mM) and 190 µL distilled water. Dilute standards in 1.5- mL centrifuge tubes as described in the Table. Transfer 20 µL Standards into separate wells of a clear flat bottom 96-well plate.

No	Premix + H ₂ O	Isocitrate (µM)
1	100 µL + 0 µL	5000
2	60 µL + 40 µL	3000
3	30 µL + 70 µL	1500
4	0 µL + 100 µL	0

- Transfer 20 µL of each sample into separate wells.
- Prepare enough Working Reagent (WR) for all assay wells by mixing, for each well, 8 µL NADP/MTT Solution, 1 µL Enzyme A, 1 µL Enzyme B, and 75 µL Assay Buffer. Fresh reconstitution of the WR is recommended.
- Add 80 µL WR to each sample well. Tap plate briefly to mix.
- Incubate at room temperature for 10 min. Use a plate reader to read OD_{565nm}.

Calculations

Subtract blank value (water, #4) from the standard values and plot the ΔOD against standard concentrations. Determine the slope and calculate the Isocitrate concentration of Sample as follows

$$[\text{Isocitrate}] = \frac{\text{OD}_{\text{SAMPLE}} - \text{OD}_{\text{BLANK}}}{\text{Slope } (\mu\text{M}^{-1})} \times n \quad (\mu\text{M})$$

where OD_{SAMPLE}, OD_{BLANK} are optical density values of the Sample and H₂O Blank, respectively. n is the sample dilution factor.

Note: if the calculated concentration is higher than 5000 µM, dilute sample in water and repeat assay. Multiple the result by the dilution factor.

Unit conversion: 1 µM is equiv. to 189 µg/L or 0.189 ppm isocitrate.

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121


www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com

Isocitrate Assay Kit (Colorimetric)

LS-K218-100 (100 Tests) • Store at -20°C



Sample Data



Isocitrate Standard Curve

Version: V.08.09.2018

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121
www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com