

## Directions for Use

### Enteral Syringe

ENFit®

Rx Only

Not made from natural rubber latex.

Not made with Diethylhexylphthalate (DEHP).

#### **WARNING:**

- The ENFit syringe is not intended for intravascular or intravenous use.
- This device has the potential to misconnect with small bore connectors of other healthcare applications.

#### **CAUTIONS:**

- Contents are sterile in an unopened, undamaged package. Do not use this device if opened or damaged.
- Do not re-sterilize.
- The patient's age, condition, and ability to feed orally will determine the method of feeding.

#### **INDICATIONS FOR USE:**

- The ENFit Syringe is intended to be attached to a nasogastric feeding tube and/or extension set for introduction of enteral fluids and/or oral medication. The ENFit Syringe is disposable and single use only.

#### **RECOMMENDED PROCEDURE:**

- Use clean technique.
- Ensure this device is only connected to an enteral port and not to an I.V. set.
- Attach ENFit female connector to ENFit male connector of feeding tube or extension set. Check to ensure the syringe connector is fully engaged to the feeding tube or extension set.
- Secure syringe to syringe pump.

#### **SYRINGE REMOVAL:**

- Remove the syringe and discard per hospital protocol.